

LA-UR-19-26512

Approved for public release; distribution is unlimited.

Title: Los Alamos National Laboratory Meteorology Monitoring Program: 2018
Data Completeness/Quality Report

Author(s): Bruggeman, David Alan
Stanton, Gregory T.
Coronado, Melissa A.

Intended for: Report

Issued: 2019-07-10

Disclaimer:

Los Alamos National Laboratory, an affirmative action/equal opportunity employer, is operated by Triad National Security, LLC for the National Nuclear Security Administration of U.S. Department of Energy under contract 89233218CNA000001. By approving this article, the publisher recognizes that the U.S. Government retains nonexclusive, royalty-free license to publish or reproduce the published form of this contribution, or to allow others to do so, for U.S. Government purposes. Los Alamos National Laboratory requests that the publisher identify this article as work performed under the auspices of the U.S. Department of Energy. Los Alamos National Laboratory strongly supports academic freedom and a researcher's right to publish; as an institution, however, the Laboratory does not endorse the viewpoint of a publication or guarantee its technical correctness.

Los Alamos National Laboratory Meteorology Monitoring Program: 2018 Data Completeness/ Quality Report

Authors: David Bruggeman, Greg Stanton, and Melissa Coronado
(Los Alamos National Laboratory Environmental Compliance Programs)

Los Alamos National Laboratory, an affirmative action/equal opportunity employer, is operated by the Los Alamos National Security, LLC, for the National Nuclear Security Administration of the U.S. Department of Energy under contract DE-AC52-06NA25396. By acceptance of this article, the publisher recognizes that the U.S. Government retains a nonexclusive, royalty-free license to publish or reproduce the published form of this contribution, or to allow others to do so, for U.S. Government purposes. Los Alamos National Laboratory requests that the publisher identify this article as work performed under the auspices of the U.S. Department of Energy. Los Alamos National Laboratory strongly supports academic freedom and a researcher's right to publish; as an institution, however, the Laboratory does not endorse the viewpoint of a publication or guarantee its technical correctness.

CONTENTS

ABSTRACT.....	1
INTRODUCTION.....	1
TA-6	3
TA-49	3
TA-53	4
TA-54	4
TA-5 MDCN.....	5
NCOM	5
REFERENCE	5

ABSTRACT

This report summarizes data completeness by tower and by instrument for 2018 and compares that data with the Los Alamos National Laboratory (LANL) and American National Standards Institute (ANSI) 2015 standards. This report is designed to make data users aware of data completeness and any data quality issues. LANL meteorology monitoring goals include 95% completeness for all measurements. The ANSI 2015 standard requires 90% completeness for all measurements. This report documents instrument/tower issues as they impact data completeness.

INTRODUCTION

LANL operates four mesa-top meteorology towers: Technical Area (TA) 06, TA-49, TA-53, and TA-54. An additional tower is located in Mortandad Canyon (TA-5 MDCN), and a rain gauge at North Community (NCOM).

A description of the meteorology monitoring network is found in Dewart and Boggs (2014). Mesa-top towers are instrumented at 1.2 meters (m), 11 m, 23 m, and 46 m. In addition, TA-06 is instrumented at 92 m. The TA-5 MDCN tower is 10 m in height and is instrumented at 1.5 m and 10 m. Data are collected every 15 minutes. Range checking is done on each measurement every 15 minutes; data that are beyond normal ranges are eliminated from the data set and replaced by a code for missing data. In addition, data are reviewed weekly by a meteorologist to identify bad data not identified by range checking. The data analyst eliminates these data from the data set and replaces them with a code for missing data. The instrument technician also reviews that data and schedules instrument replacement as required. Data completeness is determined by the number of total 15-minute records available versus the total number of possible measurements for the entire year. As a rule, LANL Environmental Compliance Programs subject matter experts do not attempt to estimate data that are eliminated as bad data. Original datalogger records, containing bad data, can be recalled from program archival storage.

The majority of missing data occur for short periods of time as a result of:

- towers down for instrument swap out/calibration
- tower hoist inspections
- power failures/network communication issues
- wind propellers freezing in snowstorms
- temperature probe aspiration fan failure
- battery failure for solar/terrestrial radiation instruments

Only other primary instrumentation failures will be documented in this report.

Summary of 2018 data completeness for all stations to meet 90% annual completeness

Sensor	TA-06	TA-49	TA-53	TA-54	TA-05	NCOM
Wind Speed Level 1	✓	✓	✓	✓	✓	
Wind Speed Level 2	✓	✓	✓	✓		
Wind Speed Level 3	✓	✓	✓	✓		
Wind Speed Level 4	✓					
Wind Direction Level 1	✓	✓	✓	✓	✓	
Wind Direction Level 2	✓	✓	✓	✓		
Wind Direction Level 3	✓	✓	✓	✓		
Wind Direction Level 4	✓					
Vertical Wind Speed Level 1	✓	✓	✓	✓	✓	
Vertical Wind Speed Level 2	✓	✓	✓	✓		
Vertical Wind Speed Level 3	✓	✓	✓	✓		
Vertical Wind Speed Level 4	✓					
Temperature Level 0	✓	✓	✓	✓	✓	
Temperature Level 1	✓	✗	✓	✓	✓	
Temperature Level 2	✓	✓	✓	✓		
Temperature Level 3	✓	✓	✓	✓		
Temperature Level 4	✓					
Pressure Level 0	✓			✗		
Relative Humidity Level 0	✓	✓	✓	✓	✓	
Dew Point Temperature Level 0	✓	✓	✓	✓	✓	
Precipitation Level 0	✓	✓	✓	✓		✓
Snow Depth Level 0	✓					
Shortwave (incoming) Level 0	✓	✓	✓	✓	✓	
Shortwave (reflected) Level 0	✓			✓		
Longwave (incoming) Level 0	✓			✓		
Longwave (outgoing) Level 0	✓			✓		
Fuel Moisture Level 0	✓					

TA-6

Summary of TA-6 2018 data completeness

Sensor	Tower Levels	Percentage Complete
Wind Speed	1 – 4	>99%
Wind Direction	1 – 4	>99%
Vertical Wind Speed	1 – 4	97%
Temperature	0 – 4	>99%
Pressure	0	>99%
Relative Humidity	0	>99%
Dew Point Temperature	0	>99%
Precipitation	0	>99%
Snow Depth	0	>99%
Shortwave (incoming)	0	>99%
Shortwave (reflected)	0	>99%
Longwave (incoming)	0	>99%
Longwave (outgoing)	0	>99%
Fuel Moisture	0	>99%

All TA-6 tower instruments exceeded the 90% data completeness in 2018.

TA-49

Summary of TA-49 2018 data completeness

Sensor	Tower Levels	Percentage Complete
Wind Speed	1	93%
Wind Speed	2, 3	>99%
Wind Direction	1	93%
Wind Direction	2, 3	>99%
Vertical Wind Speed	1	93%
Vertical Wind Speed	2, 3	>99%
Temperature	0	>99%
Temperature	1	82%
Temperature	2, 3	>99%
Relative Humidity	0	96%
Dew Point Temperature	0	>99%
Precipitation	0	>99%
Shortwave (incoming)	0	>99%

All TA-49 tower instruments exceeded the 90% data completeness in 2018, except for the temperature instrument at level 1 (82%) as a result of cable and conduit issues on the tower.

TA-53

Summary of TA-53 2018 data completeness

Sensor	Tower Levels	Percentage Complete
Wind Speed	1 – 3	>99%
Wind Direction	1 – 3	>99%
Vertical Wind Speed	1 – 3	98%
Temperature	0 – 3	>99%
Relative Humidity	0	>99%
Dew Point Temperature	0	>99%
Precipitation	0	>99%
Shortwave (incoming)	0	>99%

All TA-53 tower instruments exceeded the 90% data completeness in 2018.

TA-54

Summary of TA-54 2018 data completeness

Sensor	Tower Levels	Percentage Complete
Wind Speed	1 – 3	>99%
Wind Direction	1 – 3	>99%
Vertical Wind Speed	1 – 3	98%
Temperature	0 – 3	>99%
Pressure	0	83%
Relative Humidity	0	>99%
Dew Point Temperature	0	>99%
Precipitation	0	>99%
Shortwave (incoming)	0	>99%
Shortwave (reflected)	0	>99%
Longwave (incoming)	0	>99%
Longwave (outgoing)	0	>99%

All TA-54 tower instruments exceeded the 90% data completeness in 2018, except for the pressure instrument at level 0 (83%) had an issue from May through July. The instrument was replaced and the issue was corrected.

TA-5 MDCN

Summary of TA-5 MDCN 2018 data completeness

Sensor	Tower Levels	Percentage Complete
Wind Speed	1	>99%
Wind Direction	1	>99%
Vertical Wind Speed	1	>99%
Temperature	0, 1	>99%
Shortwave (incoming)	0	>99%

All TA-5 MDCN tower instruments exceeded 99% data completeness in 2018.

NCOM

Sensor	Tower Levels	Percentage Complete
Precipitation	0	>99%

The NCOM rain gauge exceeded 99% data completeness for 2018.

REFERENCE

Dewart, J. and M. Boggs 2014: "Meteorological Monitoring at Los Alamos," Los Alamos National Laboratory report LA-UR-14-23378.